

Product change notification

Products affected:  **ORACollect®**

Part Numbers: OC-100

Description: ORACollect® oral sample collection kit

Effective date of change: May 2013

Customer action: Only required if OC-100 is integrated with automation systems.

Change description:

Based upon user feedback, DNA Genotek has made improvements to our OC-100 product. We will be changing the double-ended cap and the collection tube to the following dimensions:

Dimensions (mm)	New	Existing
Cap diameter	18.1 mm	20.1 mm
Collection tube height	92.1 mm	96.6 mm
Depth of tube, from top of tube to bottom of false-bottom	52.1 mm	52.3 mm
Depth of tube, from bottom of tube to bottom of false-bottom	40.0 mm	44.3 mm
Outside diameter (OD) of tube	15.0 mm	15.5 mm
Inside diameter (ID) of tube at threaded end	13.0 mm	13.4 mm
Inside diameter (ID) of tube at bottom of tube	13.5 mm	12.6 mm

Reason for change:

Reduced cap and tube dimensions to improve kit handling and compatibility with liquid automation platforms.

Note:

We recommend that automated liquid-handling protocols be updated to reflect the new tube dimensions, but expect that no other re-validation work will be required as a result of this enhancement.

Please contact info@dnagenotek.com if you have any issues or concerns pertaining to this product change notification.