

Product change notification

Products affected: OMNIGENE-ORAL (tongue), OMNIGENE-ORAL (gum/plaque),
OMNIGENE-VAGINAL, OMNIGENE-SKIN

Catalog numbers/SKUs: OMR-110, OMR-120, OMR-130, OMR-140

Effective date of change: December 2024

Change description:

As part of the continuous improvement process at DNA Genotek, the shelf life of the OMNIgene™•OMR-1xx devices has been extended from 15 to 24 months. This change will allow customers increased time to use their OMNIgene™• OMR-1XX devices.

This new extended shelf life will impact all OMNIgene[™]• OMR-1XX devices, including OMNIgene[™]• ORAL (tongue), OMNIgene[™]• ORAL (gum/plaque), OMNIgene[™]• VAGINAL and OMNIgene[™]• SKIN devices.

If the current expiration on the tube label reflects an expiry within the year 2025, the product may be stored for an additional 9 months from the written use by/expiry date. The tube label of the newly manufactured product with expiry after the year 2025 will reflect the extended shelf life from 15 months to 24 months.

Customer action:

This change to the use by dates is retroactive for all products.

See the table for examples of existing use by dates and their extended use by dates:

Use by date on tube label (Year-month-date)	Extended use by date (Year-month-date)
2025-12-18	2026-09-18
2025-10-23	2026-07-23
2025-07-02	2026-04-02
2025-05-09	2026-02-09
2025-03-18	2025-12-18
2025-02-16	2025-11-16

Contact your account manager or info@dnagenotek.com if you have any questions or want to discuss this product change notification.

